

Titel **Pelikloon anti-A monoclonal**

K1188

Clone number: BIRMA1



Analysis	Requirement	Result	
Specificity	Negative with cells negative for the A antigen	Conform requirements	
Avidity		A ₁ A ₂ A ₂ B	
	Tube method	≥ 3+ ≥ 3+ ≥ 3+	Conform requirements
	Microplate method	≥ 3+ ≥ 3+ ≥ 3+	Conform requirements
	Cellbind Screen method	≥ 3+ ≥ 3+ ≥ 3+	Conform requirements
	Cellbind Direct Type method	≥ 3+ ≥ 3+ ≥ 3+	Conform requirements
Titer		A ₁ A ₂ B	
	Tube method	≥ 1:256 ≥ 1:64	Conform requirements
	Microplate method	≥ 1:256 ≥ 1:64	Conform requirements
Growthtest	no growth	Conform	
External release	released by Notified Body	Released Y/N	

Date

: 28 July 2022

Approved by QA/QC



VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 01 juli 2022 at: 12:15:31 pm ; Printed by M.BOUWMAN