


Titel **Pelikloon anti-A,B (IgM) monoclonal release**


K 1190

Clone numbers: ES4/ES15



Analysis	Requirement				Result
Specificity	Negative with O cells				Conform requirements
Avidity		A <sub>1</sub>	A <sub>2</sub>	B	
	Tube method	≥ 3+	≥ 3+	≥ 3+	Conform requirements
	Microplate method	≥ 3+	≥ 3+	≥ 3+	Conform requirements
	Cellbind Screen method	≥ 3+	≥ 3+	≥ 3+	Conform requirements
	Cellbind Direct Type method	≥ 3+	≥ 3+	≥ 3+	Conform requirements
Titer		A <sub>1</sub>	A <sub>2</sub>	B	
	Tube method	≥ 1:256	≥ 1:64	≥ 1:128	Conform requirements
	Microplate method	≥ 1:256	≥ 1:64	≥ 1:128	Conform requirements
Growthtest	no growth				Conform
External release	released by Notified Body				Released <input checked="" type="checkbox"/> Y / <input type="checkbox"/> N

Date : 31 October 2022 Approved by QA/QC



VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 18 okt 2022 at: 02:54:55 pm ; Printed by M.BOUWMAN