

Titel **Pelikloon anti-Jka (IgM) monoclonal release**

K 1300

Clone nr. MS15



Analysis	Requirement	Result
Specificity	Tube method Negative with cells negative for the Jk <sup>a</sup> antigen	Conform requirements
Avidity	Tube method	
	Jk(a+b-) ≥ 2+ Jk(a+b+) ≥ 2+	Conform requirements Conform requirements
Titer	Tube method	Conform requirements
	Jk(a+b+) ≥ 1:8	
Growthtest	no growth	Conform

Date : 18 July 2022 Approved by QA/QC : [Signature]

VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 15 juli 2022 at: 02:23:38 pm ; Printed by M.BOUWMAN