

Titel **MASPAT kit release**

K 1360

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

MASPAT kit

Monoclonal Antibody Solid Phase
Platelet Antibody Test

IVD  96

REF K1360

LOT 8000458140



2-8 °C

2024-05-02

(01)1872068203
0025(10)800045
8140(17)240502

Component	Requirement	Result
Positive control Lot: 8000456733	Positive reaction with: platelets Titration with: platelets $\geq 1:4$	Conform requirements Conform requirements
Negative control Lot: 8000456734	Negative reaction with: platelets	Conform requirements
Anti-IgG Lot: 8000456731	Titration with: platelets $\geq 1:4$	Conform requirements
Microplates Lot: 8000457067	Specificity: Platelets x positive control positive Platelets x negative control negative Coating tested platelets for reactivity and homogeneity	Conform requirements Conform requirements Conform requirements
LISS Lot: 8000457170	Conductivity 3.4 – 4.8 mScm ⁻¹ pH 6.3 – 6.9	Conform requirements Conform requirements
Growth test	no growth	Conform


All the sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with the current EU guidelines to GMP and found to be non reactive

Date

: 01 December 2022

Approved by QA/QC

:



VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 18 nov 2022 at: 12:16:52 pm ; Printed by M.BOUWMAN