


Titel **MASPAT kit release**

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

K 1360

**MASPAT kit**  
 Monoclonal Antibody Solid Phase  
 Platelet Antibody Test



IVD  96  
 REF K1360  
 LOT 8000458393


2-8 °C  
 2024-08-10

(01)1872068203  
 0025(10)800045  
 8393(17)240810

Component	Requirement	Result
Positive control Lot: 8000457174	Positive reaction with: platelets Titration with: platelets $\geq$ 1:4	Conform requirements Conform requirements
Negative control Lot: 8000457175	Negative reaction with: platelets	Conform requirements
Anti-IgG Lot: 8000457311	Titration with: platelets $\geq$ 1:4	Conform requirements
Microplates Lot: 8000457339	Specificity: Platelets x positive control positive Platelets x negative control negative Coating tested platelets for reactivity and homogeneity	Conform requirements Conform requirements Conform requirements
LISS Lot: 8000457614	Conductivity 3.4 – 4.8 mScm <sup>-1</sup> pH 6.3 – 6.9	Conform requirements Conform requirements
Growth test	no growth	Conform

All the sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with the current EU guidelines to GMP and found to be non reactive

Date : 10 January 2023 Approved by QA/QC :



VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 19 dec 2022 at: 08:30:37 am ; Printed by M.BOUWMAN