

Titel **MASPAT kit release**

K 1360

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

**MASPAT kit**  
 Monoclonal Antibody Solid Phase  
 Platelet Antibody Test



REF K1360

LOT 8000458720




2-8 °C

2024-08-10

(01)1872068203  
 0025(10)800045  
 8720(17)240810

Component	Requirement	Result
Positive control Lot: 8000457174	Positive reaction with: platelets Titration with: platelets $\geq$ 1:4	Conform requirements Conform requirements
Negative control Lot: 8000457175	Negative reaction with: platelets	Conform requirements
Anti-IgG Lot: 8000457311	Titration with: platelets $\geq$ 1:4	Conform requirements
Microplates Lot: 8000457426	Specificity: Platelets x positive control    positive Platelets x negative control    negative Coating tested platelets for reactivity and homogeneity	Conform requirements Conform requirements Conform requirements
LISS Lot: 8000457614	Conductivity 3.4 – 4.8 mScm <sup>-1</sup> pH 6.3 – 6.9	Conform requirements Conform requirements
Growth test	no growth	Conform

All the sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with the current EU guidelines to GMP and found to be non reactive

Date : 02 February 2023 Approved by QA/QC : 

VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 25 Jan 2023 at: 08:47:26 am ; Printed by M.BOUWMAN