

Titel **MASPAT kit release**

K 1360

Sanquin Plesmanlaan 125 1066 CX Amsterdam NL

**MASPAT kit**  
**Monoclonal Antibody Solid Phase**  
**Platelet Antibody Test**



IVD  96

REF K1360

LOT 8000458833




2-8 °C

2024-08-16

(01)1872068203  
 0025(10)800045  
 8833(17)240816

Component	Requirement	Result
Positive control Lot: 800045851g	Positive reaction with: platelets Titration with: platelets $\geq$ 1:4	Conform requirements  Conform requirements
Negative control Lot: 8000457175	Negative reaction with: platelets	Conform requirements
Anti-IgG Lot: 8000457311	Titration with: platelets $\geq$ 1:4	Conform requirements
Microplates Lot: 8000457541	Specificity: Platelets x positive control positive Platelets x negative control negative Coating tested platelets for reactivity and homogeneity	Conform requirements Conform requirements Conform requirements
LISS Lot: 8000457614	Conductivity 3.4 – 4.8 mScm <sup>-1</sup> pH 6.3 – 6.9	Conform requirements Conform requirements
Growth test	no growth	Conform

All the sera used in the production of this reagent have been tested for the markers of specific disease transmitting agents in accordance with the current EU guidelines to GMP and found to be non reactive

Date : 28 February 2023 Approved by QA/QC : 

VERTROUWELIJK