



Doc. nr. REA-S003-00-06-0500

Versie 2.0


Datum 06 May 2021 13:23:06 (GMT+01:00)

Pagina 1 van 2

Document eigenaar BBR

Titel Cellbind Screen release

K 7000



| Analysis | Requirement | Result |
|-----------------|---|----------------------|
| Specificity | Negative reaction with A, B and O cells | Conform requirements |
| | Negative reaction with Cellbind P3 cells tested with serum and plasma from donors and plasma from pregnant women | Conform requirements |
| Avidity / Titer | Reaction with Coombs Control Cells tested with AB serum is at least 3+ | Conform requirements |
| | Results with anti- Fy ^a , anti- Fy ^b and anti-s are at least comparable with reference method | Conform requirements |

Date : 01 December 2022 Approved by QA/QC

VERTROUWELIJK

Use copy according to REA-SOP-0002.

Printed on: 18 okt 2022 at: 03:08:03 pm ; Printed by M.BOUWMAN